

शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, गोंदिया
GOVERNMENT MEDICAL COLLEGE & HOSPITAL, GONDIA

Ref. No./GMCG/PS/Qt/

२२०४

/२०२०

Date 26/09/2020

Open Quotation Notice

For the supply of FULLY AUTOMATIC OXYGEN CONTROL PANEL

Dean Govt Medical college, Gondia, invites open quotation for the supply of Chemical for various Department from Manufacturer/Dealers. Details list & specification is enclosed. Interested supplier/dealer can send their quotation.


Quotation should be sent in sealed envelope (duly sealed by selling wax) by post or hand delivery so as to reach this office on or before 04/12/2020 12:00 upto Date & timing of the opening of quotation 05/12/2020 at 12pm at college council Hall. Representative of the supplier may attend the meeting by submitting authorization letter.

1. The "Quotation" for OXYGEN MANIFOLD & OXYGEN CONTROL PANEL FULLY AUTOMATIC and last Date / / should be clearly written on cover.
2. Quotation, which received late, will not be accepted under any circumstances.
3. Rate should be for free delivery at Govt. Medical college & Hospital, Gondia premises only. Rate should be quoted including all charges. Rates must be quote on letter head along with Pan, VAT, CST Number.
4. Quoted goods should be strictly accordingly to the specification mentioned in the list. Make Model of items should be specifically stated in quotation and catalogue/leaflets etc. should be submit along with the quotation.
5. After placing the order to lowest one, the order will have to execute in full within the stipulated time if the supplier fail to comply, within stipulated period after giving the order may be cancelled and he will be declared defaulter and his any offer will not be consider in future.
6. The undersigned reserved the rights to accept or reject any or all quotation without giving any reason.
7. The quotation should be sent in the name of the Dean, Govt. Medical College & Hospital, Gondia.
8. Payment of the order goods will be made within 1 months from the date of supply & installation of the goods by CMP. For the CMP purpose (unregistered) supplier should submit copy of Pan Card, Cancelled Check, One photograph, & Bank Details.
09. All inclusive warranty period should be two years from the date of commissioning of equipments.

Terms and conditions

- 1) Quotations are invited in Two bid envelop system. **one Envelope No. 1** contains authority letter, specifications compliance chart, brochures and catalog .
- 2) **Second Envelope No. 2** contains only price of equipment/instrument.

Envelope No. 2




Dean
Government Medical college,
Gondia

शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, गोंदिया
GOVERNMENT MEDICAL COLLEGE & HOSPITAL, GONDIA

Detail Of Work

अ. क्र.	नाव	Specification	संख्या	किंमत
०१	FULLY AUTOMATIC OXYGEN CONTROL PANEL	a) The control panel cover is made using high quality MS duly powder coated. b) The supply of gas does not depend on electricity c) The manifold shall be a ,fully automatic type and shall permit automatic switch over for the "in use" (primary) bank of cylinders to the "Reserve" (secondary) bank of cylinders d) The manifold shall have a microprocessor based display panel. The unit will be compact measuring 22"x28" wide x 6" deep. Fully automatic duly tested for hours before delivery. e) The control panel incorporate twelve large LED Red /yellow / green led display for the left bank the right bank and for delivery pressure. f) The manifold shall be equipped with a 3 /4" outlet shutoff valve. valve " parker Hannifin (Italy)CE marked "Equipped with internal power supply module 12 V-2 Amp. SMPS adaptor . low pressure regulator is fully pneumatic which avoid diaphragm leakage	१ संच	



Dean

Government Medical college & Hospital
Gondia