

शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, गोंदिया

GOVERNMENT MEDICAL COLLEGE & HOSPITAL, GONDIA

Ref. No./GMCG/PS/Qt/

1886

/2023

Date: 24/3/2023

Quotation Notice

For the supply of Instrument, Materials & Equipment


Dean Govt Medical college, Gondia, invites open quotation for the supply of Equipment for various Department from Manufacturer/Dealers. Details list & specification is enclosed. Interested supplier/dealer can send their quotation.

Quotation should be sent in sealed envelope (duly sealed by selling wax) by post or hand delivery so as to reach this office on or before 28/3/23 upto 3 Date & timing of the opening of quotation 28/3/23 at 5 pm at college council Hall. Representative of the supplier may attend the meeting by submitting authorization letter.

1. The "Quotation" for Materials & Equipment and last Date 28/3/23 should be clearly written on cover. upto 3 PM
2. Quotation, which received late, will not be accepted under any circumstances.
3. Rate should be for free delivery at Govt. Medical college, Gondia premises only. Rate should be quoted including all charges. Rates must be quote on letter head along with Pan, VAT, CST Number.
4. Quoted goods should be strictly accordingly to the specification mentioned in the list. Make Model of items should be specifically stated in quotation and catalogue/leaflets etc. should be submit along with the quotation.
5. After placing the order to lowest one, the order will have to executer in full within the stipulated time if the supplier fail to comply, within stipulated period after giving the order may be cancelled and he will be declared defaulter and his any offer will not be consider in future.
6. The undersigned reserved the rights to accept or reject any or all quotation without giving any reason.
7. The quotation should be sent in the name of the Dean, Govt. Medical College, Gondia.
8. Payment of the order goods will be made within 1 months from the date of supply & installation of the goods by CMP. For the CMP purpose (unregistered) supplier should submit copy of Pan Card, Cancelled Check, One photograph, & Bank Details.
9. The supplier should deliver the Materials & Equipment in the concerned department of the institution.

Sr.No.	Item Name	Req. Qty.
1	Strik Retinoscope	1
	Direct Ophthalmoscope	1

(Specification enclosed here with)


Dean

Government Medical college & Hospital,
Gondia

Department of Ophthalmology
GOVERNMENT MEDICAL COLLEGE, GONDIA

=====


STREAK RETINOSCOPE


Specifications

- Homogenous LED illumination for precise streak image quality
- ParaStop: Precise, easy selection of a parallel beam.
- Precise streak image with a line width of typically 1.1 mm (ISO 12865 Standard < 1.5 mm) and a line length of 35 mm
- light intensity dimmability- stepless dimming from 3 % to 100 % with one-finger operation.
- Single control for vergence and rotation
- Maintenance-free. Dustproof. LED
- Integrated polarization filter.
- Detachable brow rest.
- Should be operable on two torch cells
- Case (company supply)
- Warranty card/letter- 2 years
- User manual

Quality control:

Equipment should be as per specifications and of standard quality. On demo, if equipments quality is sub-standard or poor or the performance of the equipment/ instrument is not as per specifications, the equipment / instrument will be rejected even after purchase order is given and said purchase order will be cancelled.


24.12.22
Assistant Professor
Department of Ophthalmology
Government Medical College
GONDIA


Professor & HOD
Department of Ophthalmology
Government Medical College
GONDIA

Department of Ophthalmology
GOVERNMENT MEDICAL COLLEGE, GONDIA

=====


Direct Ophthalmoscope


Specification:

- LED illumination
- Bulb-Maintenance-free for a lifetime.
- Aspherical optical system
- Aluminum frame system
- Dustproof design
- Optical components should be flush-mounted on a cast aluminum frame,
- No misalignment
- Shock-resistant.
- On wheel -Circular aperture with full circle, half circle illumination, slit, graticule scale for measuring macular lesion
- Small pupils system
- Should have blue filtre, red free filtre
- Ophthalmoscope should accomodate 28 single lenses Disc onwheel (-36D to +38D) .
- Stepless dimming from 3 % to 100 % with practical one-finger operation.
- Suitable for large as well as small or undilated pupils.
- No stray light to eliminate distractions.
- Precise one-hand control of the instrument.
- Should be operable on in-built rechargeable battery
- Inbuilt rechargeable battery
- USB handle/ plug-in charger
- One extra bulb
- Case (company supply)
- Warranty card/letter- 2 years
- User manual

Quality control:

Equipment should be as per specifications and of standard quality. On demo, if equipments quality is sub-standard or poor or the performance of the equipment/ instrument is not as per specifications, the equipment / instrument will be rejected even after purchase order is given and said purchase order will be cancelled.


Assistant Professor 21.12.22
Department of Ophthalmology
Government Medical College
GONDIA


Professor & HOD
Department of Ophthalmology
Government Medical College
GONDIA