



महाराष्ट्र शासन

शासकीय वैद्यकीय महाविद्यालय, गोंदिया

Government Medical College, Gondia

कुंवर तिलकसिंह सामान्य रुग्णालय परिसर, नेहरु चौक, गोंदिया, महाराष्ट्र-४४१६०१

Telephone No 07182-238090

Email-ID:- gmcgstore2015@gmail.com

Website: www.gmcgondia.in

Ref. No./GMCG/PS/QV

5301

/2019

Date: 7/11/2019

Open Quotation Notice

For the supply of Instrument, Materials & Equipments

Dean Govt. Medical College, Gondia invites open quotation for the supply of INSTRUMENTS & EQUIPMENTS for various Departments from Manufacturer/Dealers. Detail list is enclosed. Interested suppliers/Manufacturer/dealer can send their quotation.

Quotation should be sent in sealed envelope (duly sealed by selling wax) by post or hand delivery so as to reach this office on or before 20/11/2019 upto 4.00 pm Date & timing of the opening of quotation 20/11/2019 at pm at college council Hall. Representative of the supplier may attend the meeting by submitting authorization letter.


1. The word "Quotation" for Instruments, Materials & Equipments and last Date 20/11/2019 Should be clearly written on cover.
2. Quotation, which received late, will not be accepted under any circumstances.
3. Rate should be for free delivery at Govt. Medical College, Gondia premises only. Rate should be quoted including all charges. Rates must be quote on letter head along with Pan and GST Number.
4. Quoted goods should be strictly accordingly to the specification mentioned in the list. Make Model of items should be specifically stated in quotation and catalogue /leaflets etc. should be submit along with the quotation.
5. Catalogue/Leaflet is must quotation received and the price of the same will not be considered on leaflets/Catalogue.
6. The concern Head of the Department can ask for demo of the lowest price model and it will be binding for the supplier to give the demo in stipulated period (i.e. within 4 days) at Gondia or nearby place. If the supplier fails to give the demonstration of the same model in stipulated period then his quotation will be treated as cancelled.
7. After placing the order to lowest one, the order will have to execute in full within the stipulated time if the supplier fail to comply, within stipulated period after giving the order may be cancelled and he will be declared defaulter and his any offer will not be consider in future.
8. The undersigned reserved the rights to accept or reject any or all quotation without giving any reason.
9. The quotation should be sent in the name of the Dean, Govt. Medical College, Gondia.
10. Payment of the order goods will be made within 1 month from the date of supply & installation of the goods by CMP. For the CMP purpose (unregistered) supplier should submit copy of Pan Card, Cancelled Check, One photograph, & Bank Details.
11. All inclusive warranty period should be two years from the date of commissioning of equipments.
12. The supplier should deliver the Instruments & Equipments in the concerned department of the institution of the institution give demonstration free of cost and staff of the department if required.

DEAN

Government Medical College,
Gondia

Government Medical College, Gondia
List for quotation (State Plan (40%) 2019-20)
Name of Item with detail Specification

Sr.No	Name of Item	Rate
1	<p>LED Phototherapy Unit:-</p> <ul style="list-style-type: none"> • Should last for at least 30,000 hours. • Should be made of easily cleanable material. • Minimum irradiance should be 20 4W/cm/nm at 45 cm distance between bed and light unil. • Wave length should be 450-460 nm and should be free from UV and IR radiation. 5. Should have visual and audible alarms for the following - <ul style="list-style-type: none"> • Internal temperature >55 deger: C • If cooling fan fails • If LED current increases or decrease • If LED volt increases or decreases • Cooling fan to be provided to dissipate the heat created by LED. • Should be compact to use along with the radiant warmer and should have tilting facility so that unit is not directly under warmer • Smooth height adjustment mechanism & adjustable height of the unit • Peak spectral irradiance: at least 30 μW/cm²/nm at 30 cm distance and 50μW/cm²/nm at 20 cm distance measured with Minolta-Airshields or Olympic flux meter) • Large effective surface area: mean spectral irradiance of at least 20 μW/cm²/nm over an area of 25 an by 50 at 30cm distance. • Emissions spectrum: blue light with a peak wavelength between 440 and 470 nm. • The source of lamps (manufacturer) should be identified and a certificate of the emission spectrum government/research laboratory should be provided. • Quality control: device is produced by manufacturer with ISO 9001 certification • Electrical: device is compatible with 220-240 V50/60 HZ power input. • Durability: LEDs are protected by suitable cover made of fire resistant transparent material • Digital total Lamp time counter. • Sturdy mobile stand should be present. • At least 2 inch sturdy Castors, at least 2 should have brakes. • Height of stand's base must be less than 8 cm to fit under radiant 	


अधिकारी
शासकीय वैद्यकीय महाविद्यालय,
गोंदिया

warmers/incubators and baby ...

- Height adjustable from 1.20m to 1.50 m (measured from bottom surface of enclosure to ground:
- Light enclosure must be tiltable from horizontal to at least 30 degrees.
- One spare set of fuses per machine.
- A list of consumable items should be provided along with its price which will be frozen for the duration on

warranty and CMC


- Live demonstration of the equipment is desirable.
- Quotation must include a compliance statement and in addition, each of the above points is marked on the technical brochure. Points not covered in the brochure must be specifically addressed in a separate set of operator & service manuals with each unit.

2

ECG Machine:-

- Display 320*240 dot single colour LCD
- Input/output – output < 100 Ohm. Sensitivity 1V+ 5%, way Floating and defibrillation protection.
- Lead- Standard 12 leads, Acquisition- 12 leads simultaneously
- Patient leak current < 10 μ A
- Sampling rate 1000Hz, A/D converter – 12 bits.
- Frequency response 0.05Hz -150Hz(-3dB)
- Time constant- time constant > 3.2 s.
- CMRR > 60dB, > 100dB (Add filter)
- Enduring polarization voltage \pm 500mV.
- Noise level 15 μ Vp-p Measurement parameters
- EMG interference filter 35 Hz(-3dB)
- AC filter 50Hz/60Hz (<20dB)
- Sensitivity choice 2.3, 5, 10, 20mm/mV, error: \pm 5% standard.
- Sensitivity is 10mm/mV \pm 0.2mm/mV Wave output
- Recording way thermal printing system
- Auto record auto leads –charging.
- Power –AC mains 100V- 240VAC, 50 Hz/60Hz, 70VAC
- Battery – built in rechargeable, 14.4 V Li ion.
- Printer - thermal, manual and auto modes, Specification of recording paper 50 mm (w)x20m(i) high-
- Speed thermal paper- paper speed 25mm/s, 50mm/s, error: \pm 5%
- Standard accessories- Patient cable- 1, Chest electrode-6, Limb electrodes- 4, Power cord- 1, Cardi Jelly bottle- 1, user manual-1

5 year AMC and CMC.


अधिकाता
शासकीय वैद्यकीय महाविद्यालय,
गोंदिया.